



Calibration Station for Ventis Product Manual

Set-up
Operation



Part Number: 17153101-1
Version 2

INDUSTRIAL
SCIENTIFIC

TABLE OF CONTENTS

WARNINGS AND CAUTIONARY STATEMENTS.....	3
CAPABILITIES	4
UNPACKING THE STATION	5
USER INTERFACE.....	5
STATION PREPARATION.....	7
Hardware Overview.....	7
Software Installation and Hardware Connections.....	9
STATION USE	11
Power-on and –off.....	11
Start-up Mode	11
Idle Mode	15
Set-up Mode.....	15
Docking and Removing the Instrument	21
Calibration	23
Bump Test.....	25
SOFTWARE	25
Software Functions.....	25
Using the Software	25
DIAGNOSING COMMON PROBLEMS	36
SPECIFICATIONS	37
PERFORMANCE SPECIFICATIONS	37
WARRANTY	38
LIMITATION OF LIABILITY.....	38
CONTACT INFORMATION.....	BACK COVER

WARNINGS AND CAUTIONARY STATEMENTS

WARNING:

Read and understand this manual before operating.

WARNING:

Failure to perform certain procedures or note certain conditions may impair the performance of this product. For maximum safety and optimal performance, please read and follow the procedures and conditions listed below.

CAUTION:

For safety reasons, this equipment must be operated and serviced by qualified personnel only.

CAUTION:

Equipment is rated for indoor use only at altitudes below 2,000 m or 6,000'.

CAUTION:

Compressed gas cylinders and their contents may present specific hazards to the user. Use only in a well ventilated area. Use only in accordance with the instructions and warnings as marked on the cylinder and the appropriate Material Safety Data Sheets.

NOTE:

The station should be cleaned only with a soft cloth; do not use solvents or other liquids.

NOTE:

This equipment has been tested and found to comply with the limits for a Class A digital device, pursuant to part 15 of the FCC Rules. These limits are designed to provide reasonable protection against harmful interference when the equipment is operated in a commercial environment. This equipment generates, uses, and can radiate radio frequency energy and, if not installed and used in accordance with the instruction manual, may cause harmful interference to radio communications. Operation of this equipment in a residential area is likely to cause harmful interference in which case the user will be required to correct the interference at the user's own expense.

Contact your service representative immediately if you suspect that the station is working abnormally.

CAPABILITIES

The V•Cal™ is a stand-alone calibration station that is compatible with the Ventis™ Multi-gas Monitor. The station is available in two dedicated models, one for the aspirated instrument and one for the diffusion instrument.

The station communicates directly with a docked instrument to perform bump tests and calibrations. It can also charge an instrument equipped with a rechargeable Li-ion (lithium-ion) battery pack.

Calibration and bump test records are saved to the station’s “buffer” (or memory) which can store a total of approximately 150 records. The results for each calibration and bump test performed are automatically sent, in report form, to an external serial printer (via a TTL-serial connection) when connected.

The station can communicate with a host PC across a USB connection, when the PC is running Accessory Software. The following capabilities are included in the calibration station with respect to commands from the host PC:

- Reading/writing instrument and calibration station settings.
- Reading the instrument datalog.
- Reading the instrument event log.
- Accessing bump and calibration records from the station.

The following operating systems support Accessory Software.

- Windows XP Professional
- Windows Vista
- Windows Server 2003
- Windows Server 2008

UNPACKING THE STATION

The station's box contains the following items. Each item should be accounted for in the unpacking process.

Quantity	Part Number	Description
1	18108631	V•Cal Calibration Station (includes: one 6" or 0.1524 m tube with a white luer AND one 6" inch or 0.1524 m tube with a white luer and t-fitting)
1	17093659	Urethane tubing (4" or 0.1016 m)
1	17121310	USB cable
1	17118027	Fitting (for gas inlet)
1	17124074	Fitting (for fresh air cylinder)
1	17121070	Industrial Scientific Accessory Software Suite CD
1	17051710	Power cord
1	17142126	Power supply

Reporting a problem. After unpacking, if any item is missing or appears to have been damaged, contact a local distributor of Industrial Scientific products or Industrial Scientific Corporation. Please refer to the manual section, [Contact Information](#).

USER INTERFACE

The calibration station user interface is comprised of the following.

- Character LCD display
- Two pushbuttons
- Three LEDs

The LCD is an eight-character by two-line display. It is backlit when the station performs a task or displays the result of a task.

The station's two pushbuttons, "CALIBRATE" and "BUMP TEST", are used to initiate the performance of those functions when an instrument is docked. These buttons are also used to access the station's set-up mode where a variety of station settings can be set or changed, and where the user can access station functions.



The station's three LEDs serve to indicate the status of the station. They are also used to help communicate bump test, zero, and calibration results.

- The green LED indicates the docked instrument has passed a calibration or bump test.

- The amber LED indicates a calibration or bump test is in-progress, or that the instrument is charging.
- The red LED indicates the docked instrument has failed a calibration or bump test. The red LED is also used to indicate an error has occurred; the LCD displays the error message(s).
- If no instrument is docked, all LEDs are off, the station is in “Ready” status, and all station functions are accessible.

STATION PREPARATION

Hardware Overview



ASPIRATED MODEL

Front of station (top); Back of station (bottom).

Items 1-10 are common to both the aspirated and diffusion models.

DIFFUSION MODEL

Front of station (top); Back of station (bottom)

Items 1-10 are common to both the aspirated and diffusion models.

Diagram Number	Feature
1	Cradle lid
2	LED indicators
3	Bump test button
4	LCD display
5	Calibrate button
6	Printer connection
7	USB connection
8	Fresh air input
9	Calibration gas input
10	Power input
11*	Instrument inlet connection
12*	Aspirated connection tubing (for use when station door is <i>attached</i>)
12*	Aspirated connection tubing with T-fitting (for use when station door is <i>removed</i>)
*Aspirated station only.	

Software Installation and Hardware Connections

Refer to the above manual section, Hardware Overview to identify the parts referenced in the following instructions.

SOFTWARE INSTALLATION

1	To install the software, insert the software CD into the CD drive of the host PC. The InstallShield Wizard program automatically starts and begins the installation process. If the program does not start, open a window on the computer to view the contents of the CD; double-click on the file titled, “Setup.exe”.
2	To complete the installation, continue following the instructions as they display on the PC. Be sure to choose the desired language for the user interface. This is completed from the drop-down menu that appears on one of the first installations screens. The choices are Chinese (simplified), English (United States), French (Standard), German, or Spanish. Highlight the desired language and click the “OK” button to continue.

HARDWARE CONNECTIONS

Attaching cables and cords.

1	<i>USB cable.</i> <ul style="list-style-type: none">On the back of the station, locate the connection marked “USB”.To connect the station to the computer, insert the cable’s flat end into the computer’s USB port; plug the other end into the USB connection on the back of the station.
3	<i>Printer connection (if desired).</i> <ul style="list-style-type: none">Power-off the printer.Connect the printer cord to the connection marked “printer” on the back of the station. Tighten the captive screws to secure the connection.
3	<i>Power supply and power cord.</i> <ul style="list-style-type: none">On the back of the station, locate the connection marked, “12VDC”.Connect the power <i>supply</i> to the 12VDC connection.Connect the other end of the power supply to the <i>power cord</i>.
Connecting the gas cylinder and demand flow regulator.	
1	Attach the demand flow regulator to the gas cylinder and turn clockwise to tighten.

2	<ul style="list-style-type: none">• Connect either end of the supplied urethane tubing to the regulator's nipple; the nipple fits inside the tubing.• On the back of the station, locate the fitting marked, "CAL GAS". This is the gas intake fitting; connect the other end of the tubing to the gas intake fitting.
---	---

FOR ASPIRATED STATIONS ONLY.**Enabling the flow of calibration gas from the station to the instrument.**

The station can perform calibrations and bump tests with its door **attached** to the station (Option 1) OR **removed** from the station (Option 2). To enable the flow of calibration gas to the installed instrument, follow the instructions below for option 1 OR option 2.

Option 1: the instrument door remains attached to the station.

1	Locate the instrument inlet connection on the back of the station.
2	Locate the tubing that has a white luer at one end; the other end of the tubing has <i>no other fittings</i> . Fasten the luer to the instrument inlet connection; turn clockwise to tighten.
3	To calibrate or bump test an installed aspirated instrument, attach the other end of the tubing to the air intake nipple extending from the station's cradle lid.

Option 2: the instrument door is removed from the station.

1	Lift the door to remove it from the station; set the door aside or store for future use.
2	Locate the instrument inlet connection on the back of the station.
3	Locate the tubing that has a white luer at one end and a t-fitting at its other end. Fasten the luer to the instrument inlet connection; turn clockwise to tighten.
4	Attach the other end of the tubing directly to the instrument's air intake. To calibrate or bump test, install the instrument in the station.

STATION USE

Power-on and -off

After all hardware connections have been completed and the software has been installed, plug in the power cord to power-on the station. There is no power-on/-off switch. To power-off the station, unplug the power cord.

Start-up Mode

When initially powered on, the station warms up and performs a series of internal diagnostics. During this process, or *start-up mode*, a series of messages displays as shown below.

DISPLAY	INSTRUCTIONS
V-Cal v 3.00.07 Displays station name and software version number (e.g., 3.00.07).	No user action required.
Warming Up	No user action required.
12345678 ABCDEFG LCD Test Display	No user action required.
Verify LEDS The station's LEDs each briefly turn on, then off.	No user action required.
 Pixel Test Display	No user action required.

Checking Currents Displays as the station checks the current draw of various station subsystems to make sure they are operating correctly.	No user action required.
Checking Clock Displays as the station checks for a valid date and time on the real-time clock.	No user action required.
22Jan10 12:34:56 Displays current date and time if the clock check was successful.	No user action required.

If the station passes all diagnostics, the Ready or Charging message displays and all station functions are available.

If the station fails any diagnostic test, the red LED turns on and a system error displays on the LCD for each failure encountered. Possible errors include “Check Pump”, “Check Sol”, “Check Board”, and “Clock Error”. These error messages and all other error and status messages are described below.

- ✓ *NOTE: Cycle the power to the station to clear errors.*

STATUS MESSAGES	
DISPLAY	STATION STATUS
Waiting to...	The bump test and calibrate functions are available. Set-up mode is accessible. USB communications are accepted and processed.
Connect Alternately display when an instrument is docked as the station attempts to communicate with the instrument.	If communication with the instrument is established within three minutes, one of two messages displays depending on the status of the battery: Charging or Ready. If communication is <i>not</i> established within three minutes, the red LED turns on and this error message displays: Inst Comm Error.

Busy Wait Displays when the USB is downloading data or communicating with the station.	The bump test and calibrate functions are <i>not</i> available. Set-up mode is <i>not</i> accessible.
Charging Displays to indicate the station is charging an instrument equipped with a Li-ion battery. The amber LED turns on. <i>NOTE: Always refer to the instrument's battery icon to assess the level of charge.</i>	The bump test and calibrate functions are available. Set-up mode is accessible. USB communications are accepted and processed.
Printing ... Displays while records print.	The bump test and calibrate functions are <i>not</i> available. Set-up mode is <i>not</i> accessible. USB communications are <i>not</i> accepted or processed.
Ready Displays when an instrument is installed and is not charging. The green LED turns on.	The bump test and calibrate functions are available. Set-up mode is accessible. USB communications are accepted and processed.
Close Lid May display when a bump test or calibration is started. Indicates the station's cradle lid is not closed.	Close the cradle lid. After the lid is closed, the station automatically restarts the calibration or bump test process.
ERROR MESSAGES	
DISPLAY	STATION STATUS
Inst Error May display during a bump test or calibration. Indicates the instrument is in a system alarm state.	The bump test or calibration is automatically aborted by the station. The display persists until the instrument is removed from the station. The bump test and calibrate functions are <i>not</i> available. Set-up mode is <i>not</i> accessible. USB communications are accepted and processed.

PRN off	--
Low temp May display when a station attempts to print. Indicates the temperature inside the station is below -15°C or 5°F.	
Inst Comm Error Displays if the bump test or calibrate process is started and the station cannot establish communication with the instrument.	The bump test or calibration is automatically aborted. Set-up mode is accessible. USB communications are accepted and processed. With another press of the bump test or calibration button, the station will again attempt the process. If unsuccessful the display persists until the instrument is removed from the station.
No Inst Detected Displays if the bump test or calibrate process is started and the station does not detect the presence of an instrument.	The bump test or calibration is automatically aborted. Set-up mode is accessible. USB communications are accepted and processed. With another press of the bump test or calibration button, the station will again check for an instrument and attempt to perform the process requested. If unsuccessful, the display persists until the instrument is removed from the station.
Check Pump Check Sol Check Board May display during the diagnostic process. Indicates the board, solenoid, or pump current falls outside the acceptable limits.	Bump test and calibration functions are <i>not</i> available. Set-up mode is accessible. USB communications are accepted and processed.

Clock Error Indicates an invalid date or clock setting on the station's real-time clock.	Bump test and calibration functions are <i>not</i> available. Set-up mode is accessible. USB communications are accepted and processed.
EEPROM Error Indicates the station cannot read and/or write to its memory.	The bump test and calibrate functions are available; however, the results may not be saved to the station. Set-up mode is accessible. USB communications are accepted and processed.
Pump Error May display after a failed zero, calibration, or bump test. Indicates a possible fault with the instrument pump, <i>not</i> the station.	Re-try calibration or check the air intake on the instrument's pump to ensure it is clear. The calibration function is available. The bump test function is <i>not</i> available. Set-up mode is accessible. USB communications are accepted and processed.
Illegal Inst May display when a bump test or calibration is started. Indicates the installed instrument is not compatible with the station.	Check that the instrument is compatible with the station; refer to the manual section, <u>Capabilities</u> . Bump test and calibration functions are <i>not</i> available. Set-up mode is accessible. USB communications are accepted and processed.

Idle Mode

When the Ready or Charging message displays, the station is in *idle mode* and all station functions are available. From idle mode, the user can enter station *set-up mode*, or can bump test or calibrate an instrument. Beginning with set-up mode, each process is outlined in the following manual sections, [Set-up Mode](#), [Calibrate](#) and [Bump Test](#).

Set-up Mode

In *set-up mode*, the user can access and perform a variety of station functions, and can initially set or subsequently change station settings.

Functions Accessible in Set-up Mode.

- Print Events (instrument-related)
- Print Buffer*
- Clear Buffer*

- Cal Days
- Set Date
- Select Language
- Check System

**Buffer is synonymous with memory. It refers to the station's saved bump test and calibration records, not the datalog of the docked instrument.*

The tasks that can be accomplished within each set-up function are described below and are presented in the order in which they are encountered by the user. Instructions for completing or bypassing each task are also provided. Generally, the **BUMP TEST** button functions as a "bypass" or "no" option; it is also used to edit values (e.g., to set the date). The **CALIBRATE** button generally functions as a "begin" or "yes" option; it is also used to set values.

- ✓ *NOTE: In set-up mode, when no button is pressed within ten seconds, the screen times out and the station returns to idle mode.*

DISPLAY MESSAGE	INSTRUCTIONS
Ready or Charging	Simultaneously press BUMP TEST and CALIBRATE , hold for three seconds, and release to enter set-up mode.
Print Events When selected, all event data for the installed instrument (not the station) are sent to the printer.	Press BUMP TEST to bypass the print events function and advance to the Print Buffer function. Press CALIBRATE to begin printing.
Print OK Y N Displays to allow the user to continue with or cancel the print command. If confirmed, all event data from the installed instrument is printed.	Press BUMP TEST to bypass printing and advance to the Print Buffer function. Press CALIBRATE to send the data to the printer; the Printing Events message displays.
Printing Events Displays while the event log prints.	No user action required. <i>NOTE: While the event log prints, the station ignores all button presses and USB communications.</i>

<h3>Print Buffer</h3> <p>When selected, <i>all</i> bump test and calibration records saved to the station are sent to the printer.</p>	<p>Press BUMP TEST to bypass the print buffer function and advance to the Clear Buffer function. Press CALIBRATE to begin printing.</p> <p><i>NOTE: The station can store up to 150 bump test and calibration reports. If the buffer is nearly full, it may take more than 30 minutes to complete the printing task.</i></p>
<h3>Print OK</h3> <p>Y N</p> <p>Displays to allow the user to continue with or cancel the print command. If confirmed, <i>all</i> records saved to the station are printed.</p>	<p>Press BUMP TEST to clear the print command. The user advances to the next set-up function, "Clear Buffer". Press CALIBRATE to send the data to the printer.</p>
<h3>Printing</h3> <p>...</p>	<p>Displays while the records print. No user action required.</p> <p><i>NOTE: While buffer records print, the station ignores all button presses and USB communications.</i></p> <p><i>NOTE: After printing the saved bump test and calibration reports, they are NOT automatically deleted from the station. The "clear buffer" function is used to complete that task.</i></p>
<h3>Clear Buffer</h3> <p>When selected, <i>all</i> saved bump test and calibration records are deleted from the station.</p>	<p>Press BUMP TEST to bypass the clear buffer function. The records remain in the buffer and the user advances to the next set-up function, Cal Days. Press CALIBRATE to delete all records saved to the station.</p> <p><i>NOTE: The clear buffer function is executable regardless of whether or not the buffer has been printed.</i></p>
<h3>Clear OK</h3> <p>Y N</p> <p>Displays to allow the user to continue with or cancel the clear command. If confirmed, all bump test and calibration records saved to the station are deleted.</p>	<p>Press BUMP TEST to cancel the clear buffer command. The records remain in the buffer and the user advances to the next set-up function, Cal Days. Press CALIBRATE to delete all records saved to the station.</p>

Buffer Cleared Displays to indicate the station's records have been successfully deleted.	No user action required.
Cal Days 30 Displays the current setting for the number of days between calibrations (e.g., 30).	This function allows the user to set the elapsed time between calibrations. The setting is programmed into the station as well as the instrument. Press BUMP TEST to bypass the setting process and advance to the next set-up function, Set Date. Press CALIBRATE to edit the value.
30 ^ OK Valid values: 1 – 365 days Increment: one day	Press BUMP TEST to increase the number of days, if needed; hold to speed the increment pace. Press CALIBRATE to set the value displayed.
Set Date	Press BUMP TEST to bypass the setting process and advance to the next set-up function, Select Language. Press CALIBRATE to edit the values, if needed. Each of these values can be changed and are presented to the user in this order: month, day, year, hour, and minutes.
01 Month ^ OK Displays show for each date and time values. The station uses a 24-hour format for time settings.	The first date/time value subject to change is the month (as shown). Press BUMP TEST to edit the value, if needed. Press CALIBRATE to set the value displayed. Continue to use the BUMP TEST and CALIBRATE buttons, respectively, to edit and set all date and time values.
Select Language	Press BUMP TEST to bypass the setting process and advance to the next set-up function, Check System. Language options are presented to the user in this order: English, Espanol, Francais, and Deutsch. Press CALIBRATE to edit the language selection, if needed.

English ^ OK	Press BUMP TEST to bypass the first language displayed, English. The next available language displays. Continue to press BUMP TEST until the desired language displays. Press CALIBRATE to set the language.
Check System	With the Check System function, the user can ask the station to run a full set of diagnostics. The station will cycle through all tests described in the manual section, <u>Start-up</u> . Press BUMP TEST to bypass the diagnostics process and advance to the Exit display. Press CALIBRATE to complete the system diagnostics.
Exit	Press BUMP TEST to remain in set-up mode . The user returns to the first function in set-up mode, Print Events. Press CALIBRATE to exit set-up mode and return to idle mode.

Figure 1 lists the information contained in reports that are generated from the printing processes described above, printing events and printing the buffer (calibration and bump test reports).

Report information generated from set-up mode functions.	
Print Events	Print Buffer
<ul style="list-style-type: none"> • Industrial Scientific Corp. • Name of calibration station and its software version • Date of printout • Instrument serial number • Instrument software version • Instrument hardware version • For each alarm event: <ul style="list-style-type: none"> ◦ Sensor type ◦ Sensor serial number ◦ Sensor high alarm threshold ◦ Sensor low alarm threshold ◦ Peak gas exposure value during alarm ◦ Duration of alarm event in seconds ◦ Time and date the alarm occurred ◦ Instrument user setting ◦ Instrument site setting • A blank for the user's signature • A blank for the user to enter the date • A blank for the user to enter the time 	<ul style="list-style-type: none"> • Industrial Scientific Corp. • Name of calibration station and its software version • Date and time of calibration (or bump test) • Instrument serial number • Instrument software version • Instrument hardware version • Instrument zero, calibration, or bump test result (pass or fail) • Recommended date for next calibration (shown for calibration only; date is blank for a failed calibration) • For each sensor: <ul style="list-style-type: none"> ◦ Sensor type ◦ Span reserve (for calibration) or Final bump test reading (for bump test) ◦ Calibration gas concentration ◦ High alarm threshold ◦ Low alarm threshold ◦ Zero, calibration, or bump test results • A blank for the user to enter the cylinder lot number • A blank for the user's signature
Figure 1a. Event report.	Figure 1b. Bump test and calibration report.

Docking and Removing the Instrument

STEP INSTRUCTIONS	
Docking the ASPIRATED instrument.	
1	When the station faces the user, its cradle lid hinge is to the user's right. Lift the lid from the left to open the station and reveal its cradle.
2	To properly place the instrument in the station cradle, complete or observe the following. <ul style="list-style-type: none"> • The instrument's display faces the user and its logo is readable. • Position the instrument so its battery contacts will touch station's charging contacts. • Press down on the instrument to secure it in the cradle; the cradle's spring supports will depress. • If needed, slide the instrument forward to secure.
3	Close the cradle lid.
Installing the DIFFUSION instrument	
1	When the station faces the user, its cradle lid hinge is at the top of the station. Lift the lid from the bottom to open the station and reveal its cradle.
2	To properly place the instrument in the station cradle, complete or observe the following. <ul style="list-style-type: none"> • The instrument's display faces the user and its logo is readable. • Position the instrument so its battery contacts will touch station's charging contacts. • Press down on the instrument to secure it in the cradle; the cradle's spring supports will depress. • If needed, slide the instrument forward to secure.
3	Close the cradle lid.
Removing the DIFFUSION or ASPIRATED instrument	
1	Lift the cradle lid (as instructed above for an aspirated or diffusion instrument).
2	Lift the instrument to remove it from the station.
3	Close the cradle lid.

Calibration

DISPLAY	INSTRUCTIONS
Ready Or Charging The station checks to see if the instrument was just calibrated.	<p><i>NOTE: If a printout of the calibration report is desired, ensure the printer is on and not in fault before beginning the calibration process. Printer faults (e.g., out of paper) are not indicated on the station. If the printer is in fault, the report does not print.</i></p> <p>Press CALIBRATE to begin; hold for three seconds and release.</p>
Cal Again? Displays if the instrument has just been calibrated.	<p>Press CALIBRATE to begin the recalibration of the instrument. Press BUMP TEST to cancel the recalibration. The station returns to its ready state in idle mode.</p> <p><i>NOTE: if no button is pressed, the instrument is not recalibrated and the "Cal Pass" message displays.</i></p>
Warming up Ventis Alternately display if the instrument is charging. This indicates the station is preparing the instrument for calibration.	<p>No user action required.</p> <p>If the instrument's <i>only</i> installed sensor is O2, the Cal In Progress message displays next. For all other installed sensor combinations, the Zero in Progress message displays.</p>
Zero in Progress Displays during the zero process which requires approximately fifteen seconds to complete. The station's amber LED turns on.	<p>No user action required.</p> <p>If the zero process <i>passes</i>, the calibration process is started by the station and the Cal In-Progress message displays.</p> <p>If the zero process <i>fails</i>, the calibration is canceled and the "Cal Fail" message displays.</p>

<p>Cal in Progress</p> <p>Displays during the calibration process. If the instrument is set for standard calibration*, the gas name abbreviation and calibration gas value for each sensor display as that sensor is calibrated. The station's amber LED turns on.</p> <p>*When an instrument is configured for standard calibration, each sensor is calibrated independently, or one sensor at a time. For a quick calibration, all installed sensors are calibrated or bump tested simultaneously.</p>	<p>No user action required. The Cal Pass or a Cal Fail message displays to indicate the calibration results.</p>
<p>Cal Pass</p> <p>Displays to indicate the instrument has passed the zero and calibration processes. The station's green LED turns on. The calibration report is sent out the RS232 port for printing and is saved to the station's buffer.</p>	<p>Remove the instrument from the station to clear the display.</p>
<p>Cal Fail Or Zero Fail</p> <p>Displays to indicate which process the instrument failed. The station's red LED turns on.</p> <p>The fail message displays alternately with error messages. A pump error message indicates a problem with the instrument's pump. Sensor name message(s) indicates which sensor(s) failed the process.</p>	<p>Remove the instrument from the station to clear the display.</p> <p><i>NOTE: The instrument must pass a zero and/or calibration before it is useable.</i></p>

Bump Test

DISPLAY	INSTRUCTIONS
Ready Or Charging	<i>NOTE: If a printout of the bump test report is desired, ensure the printer is on and not in fault before beginning the bump test process. Printer faults (e.g., out of paper) are not indicated on the station. If the printer is in fault, the report does not print.</i>
Bump in Progress Displays during the bump test process. The gas name abbreviation and calibration gas value for each sensor display as that sensor is bump tested. The station's amber LED turns on.	No user action required. The station reads and applies the bump test parameters (gas percentage and response time settings) from the instrument. If one or more installed sensors are in a calibration or zero fail state, the station automatically cancels the bump test and performs a calibration.
Bump Pass Displays to indicate the instrument passed a bump test. The station's green LED turns on. The bump test report is sent out the RS232 port for printing and is saved to the station's buffer.	Remove the instrument from the station to clear the display.
Bump Fail Displays to indicate the instrument failed a bump test. The gas name abbreviation for each failed sensor also displays. The station's red LED turns on.	Remove the instrument from the station to clear the display.

SOFTWARE USE

Software functions

Accessory Software functions are organized into categories, and are presented on the software's user interface as "tabs". The tabs are listed below with descriptions of the functions accessible from each.

TAB	FUNCTION/CONTENTS
General	Administration information for the instrument.
Options	Instrument configuration options.
Users and Sites	Shows active user and site saved in instrument (not viewable on instrument).
Components	Shows details of the instrument's components.
Calibrations	Shows calibration data associated with each instrument (can view saved records or download the latest).
Bump Tests	Shows bump test data associated with each instrument (can view saved records or download the latest).
Event Log	Shows log files and associated data for each instrument.
DataLogging	Shows datalog files and associated data for each instrument.

Beginning with the General tab, each tab is reproduced in the following pages. Command icons (or buttons) also appear on each tab and accomplish the following when selected by the user.

Refresh: generally used when a new instrument is docked to access its datalog, event log, settings, etc.

Update: after editing any value on a tab, the instrument settings are updated to reflect the new value(s).

Print: opens a new window containing a printable report of the information relevant to the tab.

Disconnect: returns the user to the Connection form.

Using the software

If Accessory Software is not already running, double-click on the desktop icon to reach the Connection form. The software can also be started from “Programs” within the computer’s “Start” menu.



Figure 2. Desktop icon.




Figure 3. Connection form.

Connection options:

Complete the connection to the docked instrument. Choose the appropriate instrument name, station type, and communications port (the default is the port with the highest port number). Choose “Connect” to complete the connection with the installed instrument. The software opens to the main screen—the “General” tab—where data are editable and the download function is accessible.

Work offline. Choose “Work offline” to view previously downloaded data and reports with no device connected to the PC. The next window to open presents a list of serial numbers those of instruments available to view offline (see Figure 4). Highlight the desired serial number and click “Open”. The

software opens to the main screen—the “General” tab. When working offline, data are *not* editable and the download function is *not* accessible.




Figure 4. Offline instrument access window.

The software user can highlight an instrument serial number and select “Open” to view that instrument’s downloaded data. The accessible read-only data includes that which is associated with these tabs: General, Components, Event Log, and DataLogging, plus any downloaded calibration and bump test records.




Figure 5. General tab.

The opening software screen after connecting to a docked instrument or connecting to work offline. The fields shown in white are editable.




Figure 6. Options tab.

The instrument configuration options can be initially set and subsequently changed from this screen. A check mark indicates the option is enabled.




Figure 7. Users and Sites tab.

The software user can assign, to the docked instrument, one active user name and one active site name. This information is saved in the instrument, but not viewable on the instrument.




Figure 8. Components tab.

The components tab lists all components installed in the instrument. The software user can highlight a component and select “Open” to view and modify its settings (e.g., Sensor Details window shown below in Figure 9).




Figure 9. Sensor Details window.

The sensor details screen allows the software user to change alarm set points as well as the calibration gas concentrations for the installed sensors.




Figure 10. Calibrations tab.

Lists all certificate files for calibration results that have been downloaded for the instrument. If the software user highlights a certificate and selects the "Open File" command, that calibration certificate opens in a new window. When the "Download" command is selected, all calibration certificates are downloaded from the station.

Ventis MX4 - Calibration Certificate

Instrument S/N: 100526C-060

Calibration Date: 10/26/2010

Technician: _____

Cylinder Lot #: _____

Sensor Type	Cal Date	Span Reserve	Pass/Fail	Low Alarm	Hi Alarm
H2S	10/26/2010 2:09:00 PM	36.5	Pass	10 PPM	20 PPM
LEL	10/26/2010 2:09:00 PM	45	Pass	10 %LEL	20 %LEL
CO	10/26/2010 2:09:00 PM	208	Pass	35 PPM	70 PPM
O2	10/26/2010 2:09:00 PM	29.9	Pass	19.5 %VOL	23.5 %VOL



Figure 11. Sample Calibration Certificate.
The user can print the certificate if needed.




Figure 12. Bump Tests tab.

Lists all certificate files for bump test results that have been downloaded for this instrument. If the software user highlights a certificate and selects the "Open File" command, that bump test certificate opens in a new window. When the "Download" command is selected, all bump test certificates are downloaded from the station.

Ventis MX4 - Bump Test Certificate

Instrument S/N 100526C-060

Bump Test Date 10/26/2010

Technician: _____

Cylinder Lot #: _____

Sensor Type	Bump Date	Sensor Reading	Pass/Fail	Low Alarm	Hi Alarm
H2S	10/26/2010 12:09:00 PM	15.3 PPM	Pass	10 PPM	20 PPM
LEL	10/26/2010 12:09:00 PM	26 %LEL	Pass	10 %LEL	20 %LEL
CO	10/26/2010 12:09:00 PM	89 PPM	Pass	35 PPM	70 PPM
O2	10/26/2010 12:09:00 PM	19.1 %VOL	Pass	19.5 %VOL	23.5 %VOL



Figure 13. Sample Bump Test Certificate.
The software user can print the certificate if needed.




Figure 14. Event Log tab.

Lists all downloaded event logs for the docked instrument. If the software user highlights a log file and selects the “Open File” command, that event log report opens in a new window. When the Download command is selected, all event logs are downloaded from the connected instrument.

Ventis MX4 100526C-060

10/26/2010 11:58:59 AM
 Serial Number: 100526C-060 Access Code: 000
 Type: Ventis MX4 Calibration Interval: 33 Days
 Part Number: VTS-K123110 Recording Interval: 10 Seconds
 Job Number: 100526
 Setup Technician: MDS TWA Time base: 8 Hours
 Setup Date: 6/18/2010 User:
 Software Version: 3.00.10 Site:

INSTRUMENT OPTIONS	
Can bump in field	On
Calibration past due warning enabled	On
Alarm latching	Off
Can calibrate in field	Off
Bump past due warning enabled	On
Can zero in field	On
Can perform quick calibration	On

Sensor SN	Sensor Type	Enabled	Cal Gas	Cal Gas Conc	Low Alarm	Hi Alarm	TWA Alarm	STEL Alarm	Calibration Date
10010f7475	Hydrogen Sulfide Sensor	Yes	Hydrogen Sulfide	25 PPM	10 PPM	20 PPM	10 PPM	15 PPM	6/18/2010 9:18:33 AM
100236a046	Combustible-LEL Sensor	Yes	Pentane	25 LEL	10 LEL	20 LEL	N/A	N/A	6/18/2010 9:20:16 AM
091232w310	Carbon Monoxide Sensor	Yes	Carbon Monoxide	100 PPM	35 PPM	70 PPM	35 PPM	200 PPM	6/18/2010 9:19:28 AM
32880630060	Oxygen Sensor	Yes	Oxygen	20.9 VOL	19.5 VOL	23.5 VOL	N/A	N/A	6/18/2010 9:17:32 AM

Alarm Time	Duration	Gas	Sensor SN	Hi Alarm	Low Alarm	Peak Reading	User	Site
10/5/2010 9:21:56 AM	00:05	Oxygen	32880630060	23.5	19.5	19		
10/5/2010 9:21:41 AM	00:06	Oxygen	32880630060	23.5	19.5	18.7		

Figure 15. Sample printout for event log.

A similarly formatted report is also available for datalogs when the “Print” command is selected from the DataLogging tab.




Figure 16. DataLogging tab

The DataLogging tab lists all downloaded datalogs for the docked instrument. When the “Download” command is selected, all data logs are downloaded from the connected instrument.

When the user highlights a session, the command buttons will accomplish the following:

- The “Summary” command opens a new window with all sensor data for that session.
- The “Print” command opens in a new window that is similar in content and format, to the Event Log Report as shown above in Figure 15.
- The “Open File” command allows the software user to view the next level of detail for a highlighted session, as shown below in the Sensor Session data (Figure 17).




Figure 17. Sensor details.

By highlighting a single sensor and clicking on the "Detail" button, the user can view a complete list of readings for that sensor for that sensor session. The sensor session can be printed, shown graphically, or exported to a comma separated variable file by using the "Print", "Graph", or "Export" command buttons, respectively. The "Compare" feature allows the user to compare the sensor session data for two or more highlighted components.

DIAGNOSING COMMON PROBLEMS

Problem	Likely Cause(s)
Display is blank...	No power to the instrument; display is damaged. Contact factory.
Unit resets...	Internal error. Cycle the power. If problem persists, contact factory.
Instrument continually fails bump test or calibration...	Ensure calibration gas is connected and the bottle is full. Sensors may require replacement. Contact factory.
Printer is not working...	Ensure paper is in printer and printer ribbon is in place.
No communication to PC...	Ensure application software and the USB driver are installed on PC. Ensure USB cable is plugged in. Ensure the correct COM port is selected on the Connection window of the software.
V•Cal does not communicate with instrument...	Ensure IR ports on both the V•Cal and the instrument are clean from dirt and debris.
V•Cal PC software will not connect to instrument...	Ensure instrument is placed in the instrument cradle. Ensure IR ports on both V•Cal and instrument are clean from dirt and debris.

SPECIFICATIONS

	Aspirated	Diffusion
Instruments supported	Ventis Aspirated with Extended Range Lithium-ion (typical)	Ventis diffusion with Lithium-ion (typical)
Dimensions	172 mm x 67 mm x 66 mm (6.8" x 2.6" x 2.6")	Ventis diffusion 103 mm x 58 mm x 30 mm (4.1" x 2.3" x 1.2")
Gas Inlets	One fresh air, one gas cylinder	
Pump Flow Rate	500 ml/min	
Input	Universal AC power supply; 110 / 240 VAC, 50/60 Hz	
Communication	On-board LEDs give status indication. Multilingual LCD display shows status and set-up menus. Real-time readings on the Ventis display during calibration.	
Internal memory	Stores up to 150 bump test and calibration reports before overwrite. Memory retains information when power is off.	

PERFORMANCE SPECIFICATIONS

Category	Specification	
Operating Temperature Range	0°C to +50°C	
Storage Temperature	-20°C to +60°C	
Operating Humidity Range	0 to 80% RH up to 31°C, decreasing linearly to 50% RH at 40°C	
External Power Supply Ratings	Supply voltage	110-240 VAC
	Frequency range	50/60 Hz
	Current Rating	1.5A
Installation Category	2	
Pollution Degree	2	

Warranty

Industrial Scientific Corporation's Calibration Stations for the VENTIS are warranted to be free from defects in material and workmanship for a period of one year after purchase.

Limitation Of Liability

THE WARRANTY SET FORTH ABOVE IS STRICTLY LIMITED TO ITS TERMS AND IS IN LIEU OF ALL OTHER WARRANTIES, GUARANTEES, EXPRESS OR IMPLIED, ARISING BY OPERATION OF LAW, COURSE OF DEALING, USAGE OF TRADE OR OTHERWISE. INDUSTRIAL SCIENTIFIC MAKES NO OTHER WARRANTIES, EITHER EXPRESSED OR IMPLIED, INCLUDING BUT NOT LIMITED TO THE WARRANTIES OF MERCHANTABILITY OR FITNESS FOR PARTICULAR PURPOSE.

SHOULD THE PRODUCT FAIL TO CONFORM TO THE ABOVE WARRANTY, BUYER'S ONLY REMEDY AND INDUSTRIAL SCIENTIFIC'S ONLY OBLIGATION SHALL BE, AT INDUSTRIAL SCIENTIFIC'S SOLE OPTION, REPLACEMENT OR REPAIR OF SUCH NON-CONFORMING GOODS OR REFUND OF THE ORIGINAL PURCHASE PRICE OF THE NONCONFORMING GOODS.

IN NO EVENT WILL INDUSTRIAL SCIENTIFIC BE LIABLE FOR ANY OTHER SPECIAL, INCIDENTAL, CONSEQUENTIAL, PUNITIVE OR OTHER SIMILAR DAMAGES, INCLUDING LOSS OF PROFIT OR LOSS OF USE, ARISING OUT OF THE SALE, MANUFACTURE OR USE OF ANY PRODUCTS SOLD HEREUNDER WHETHER SUCH CLAIM IS PLEADED IN CONTRACT OR IN TORT, INCLUDING STRICT LIABILITY IN TORT AND WHETHER INDUSTRIAL SCIENTIFIC HAS BEEN ADVISED OF THE POTENTIAL FOR SUCH DAMAGES. Industrial Scientific's total liability hereunder from any cause whatsoever (except liability from personal injury caused by Industrial Scientific's negligence), whether arising under contract, warranty, tort (including negligence), strict liability, products liability or any other theory of liability, will be limited to the lesser of Buyer's actual damages or the price paid to Industrial Scientific for the Products that are the subject of Buyer's claim. All claims against Industrial Scientific must be brought within one year after the cause of action arises, and Buyer expressly waives any longer statute of limitations.

It shall be an express condition to Industrial Scientific's warranty that all products be carefully inspected for damage by Buyer upon receipt, be properly calibrated for Buyer's particular use, and be used, repaired, and maintained in strict accordance with the instructions set forth in Industrial Scientific's product literature. Repair or maintenance by non-qualified personnel will invalidate the warranty, as will the use of non-approved consumables or spare parts. As with any other sophisticated product, it is essential and a condition of Industrial Scientific's warranty that all personnel using the products be fully acquainted with their use, capabilities and limitations as set forth in the applicable product literature.

Buyer acknowledges that it alone has determined the intended purpose and suitability of the goods purchased. It is expressly agreed by the parties that any

Calibration Station for the Ventis™: Product Manual

technical or other advice given by Industrial Scientific with respect to the use of the goods or services is given without charge and at Buyer's risk; therefore, Industrial Scientific assumes no obligations or liability for the advice given or results obtained.

CONTACT INFORMATION

Industrial Scientific Corporation

1 Life Way
Pittsburgh, PA 15205-7500
USA
Web: www.indsci.com
Phone: +1 412-788-4353 or 1-800-DETECTS (338-3287)
E-mail: info@indsci.com
Fax: +1 412-788-8353

Industrial Scientific France S.A.S.

5 Rue Frédéric Degeorge, CS 80097
62002 Arras Cedex,
France
Web: www.indsci.com
Phone: +33 (0)1 57 32 92 61
E-mail: info@eu.indsci.com
Fax: +33 (0)1 57 32 92 67

英思科传感仪器（上海）有限公司

地址：中国上海市浦东金桥出口加工区桂桥路290号

邮编：201206

电话：+86 21 5899 3279

传真：+86 21 5899 3280

E-mail: info@ap.indsci.com

网址：www.indsci.com

服务热线：+86 400 820 2515

To locate a nearby distributor of our products or an Industrial Scientific service center or business office, visit us at www.indsci.com.

Rendez-vous sur notre site Web www.indsci.com, si vous voulez trouver un distributeur de nos produits près de chez vous, ou, si vous recherchez un centre de service ou un bureau Industrial Scientific.

Besuchen Sie uns unter www.indsci.com, um einen Vertriebshändler unserer Produkte oder ein Servicecenter bzw. eine Niederlassung von Industrial Scientific zu finden.

Para buscar un distribuidor local de nuestros productos o un centro de servicio u oficina comercial de Industrial Scientific, visite www.indsci.com.

如需查找就近的产品经销商或 Industrial Scientific 务中心或业务办事处, 请访问我们的网站
www.indsci.com。

INDUSTRIAL
SCIENTIFIC